



QUALITY CERTIFICATE

IBCm Bacto Kit – Component 1

P/N	Lot number	Expiration date	Instruments
10229-A	10229-A-390F	04/2019	IBCm

Parameter	Specification	Result
Appearance (color)	Clear	Clear
Appearance (form)	Liquid	Liquid
pH at 25°C	9.55 ± 0.17	9.45
Qualitative IR/FTIR analysis	Conform to structure	Conforms

Date of issue of certificate: Apr 05, 2018

Henrik Lyder, Quality Control

The reagents covered by this certificate have been produced and packaged in accordance with Bentley Instruments in-house quality control system. This is to certify that units of the lot number mentioned above were tested and found to comply with the specifications of the grade listed. The test data has been determined by our manufacturer/supplier and/or in our quality control laboratory. This analysis is not to be construed as a warranty, expressed or implied. All products from Bentley Instruments, Inc. are manufactured under an ISO 9001:2008 Quality System, Certificate ID 1004034, expiry date September 15, 2018.